

Four-year follow-up on a ZSI 375 artificial urinary sphincter for male urinary incontinence from one urological centre in Poland

[Autor's unedited version]

Ireneusz Ostrowski¹, Janusz Ciechan¹, Emil Śledź¹, Wojciech Dyś¹, Tomasz Golabek², Piotr L. Chłosta²

¹Department of Urology and Urological Oncology, Regional Specialist Hospital, Puławy, Poland

²Department of Urology, Jagiellonian University, Medical College, Cracow, Poland

Corresponding author

Ireneusz Ostrowski
irekostrowski@op.pl

Citation: Ostrowski I, Ciechan J, Śledź E, Dyś W, Golabek T, Chłosta PL. Four-year follow-up on a ZSI 375 artificial urinary sphincter for male urinary incontinence from one urological centre in Poland. Cent European J Urol. 2018; doi: 10.5173/cej.2018.1704 [Epub ahead of print]

Article history

Submitted: June 13, 2018

Accepted: June 20, 2018

Published online: June 29, 2018

Key Words: artificial urinary sphincter, radical prostatectomy, male urinary incontinence, ZSI 375

Introduction The treatment of choice for patients who have iatrogenic urinary incontinence is the implantation of an artificial urinary sphincter. We did a prospective study on the outcomes of men undergoing artificial urinary sphincter (AUS) implantation (ZSI 375; Zephyr Surgical Implants, Geneva, Switzerland).

Material and methods Patient with moderate-to-severe stress incontinence were included. The ZSI 375 device is a one-piece device with a cuff and a pump unit. Patients underwent a perineal incision for cuff insertion and an inguinal incision for pump unit to be placed in the scrotum. Complications and pads used to manage incontinence were recorded. Perioperative complications were categorized according to the Clavien-Dindo classification. Pains connected with implantation of AUS were assessed with the use of VAS (Visual Analogue Scale).

Results Between July 2013 and June 2017, 50 patients underwent ZSI 375 device insertion in the Department of Urology and Urological Oncology in Puławy. The follow-up stopped at the end of September 2017. The median (range) follow-up was 21.04 (1–50) months. No patient experienced bladder hyperactivity. Complications leading to a revision or permanent device removal arose in 12 patients (erosion = 9, infection = 0, mechanical failure = 3). Social continence (0 or 1 pad/day) was achieved in 29/50 patients (58%). An improvement (50% less pads/day) was achieved in 15/50 patients (30%). A failure was seen in 6/50 patients (12%). All perioperative patients were classified to grade I in Clavien-Dindo classification. Mean value of pain intensity in VAS was 0.82.

Cent European J Urol

Conclusions The ZSI 375 device is safe, effective and the follow-up was long enough to identify all potential complications.

INTRODUCTION

The Artificial Urinary Sphincter (AUS) was first introduced in 1973 for the treatment of male stress urinary incontinence and became the treatment of choice for male urinary incontinence [1, 2, 3]. AUS insertion remains a complexed procedure with a risk of complications such as erosion, infection and mechanical failure.

Throughout the years, there have been improvements in AUS quality, with the idea of double cuffs and new surgical approach with improved outcomes and decreased complications [3–8]. With experience and 10.5% of urethral erosion, the double cuff is now not recommended for first AUS insertion [9, 10, 11].

Recently, IZ (minocycline and rifampin) coated devices, were introduced to decrease the AUS infection rate, but without any significant success [12].

Revision rates attributable to mechanical failure still range from 8 to 45% [3, 13,14, 15].

The quality of the sphincter's preparation, components connexion, and surgeon's experience during the procedure are the key factors for success as the learning curve is long.

Most of the sphincter insertions are carried out by occasional operators (1 to 3 procedures per year). Less than 10% of AUS insertions are performed by surgeons with experience of at least 100 procedures in the US [16]. The risk of revision increases from 13 to 24% [17].

ZSI 375 (Zephyr Surgical Implants, Geneva, Switzerland) was designed to facilitate AUS insertion (Figure 1). The cuff is adjustable around the urethra (Figure 2). It is pre-connected. The pressure can be increased to improve the patient's continence. It has no abdominal reservoir so as to reduce the operating time and to avoid abdominal incision and dissection in scarred retroperitoneum [18, 19, 26].

The aim of the present study was to analyse the safety and efficacy of ZSI 375 in 50 incontinent male patients with a follow up from July 2013 to September 2017.

MATERIAL AND METHODS

Patients

We performed a prospective analysis of patients who underwent ZSI 375 placement in one urological centre in Poland from July 2013 to June 2017. Patients were followed up to September 2017. Our urological team (two surgeons) inserted the ZSI 375 artificial urinary sphincter during their learning curve period.

Indication for ZSI 375 insertion was for moderate stress urinary incontinence (2 to 3 pads/day) and severe stress incontinence (4 and more pads/day). All patients had failed previous rehabilitation by pelvic floor training and electrostimulation. The preoperative evaluation included patient history, pad use, a physical examination, and cystoscopy looking for stenosis, urine analysis and urodynamic to exclude an overactive bladder.

Device preparation and implantation

The ZSI 375 is a one-piece AUS made up of a cuff connected by a kink-resistant tubing to a pump unit (Figure 1). The adjustable cuff is moulded in a curve to avoid folding. It is placed around the urethra, the pump unit (pressure-regulating tank and pump) is placed in the scrotum, in the subdartos pouch (Figure 2). After activation, the pressure in the hydraulic circuit can be increased or decreased to improve the patient's continence.

Cent European J Urol

Two versions of the sphincter manufactured by Zephyr Surgical Implants (Geneva, Switzerland) are available: a version ZSI 375 provided dry and that must be filled with a saline solution before insertion and a pre-filled version ZSI 375 PF ready to be implanted.

The sphincters used in this prospective study were the first ZSI 375 version and the latest a pre-filled version ZSI 375 PF. The first version has to be prepared before insertion: the sphincter must be filled with a saline solution and air bubbles removed. The second is prepared before implantation.

The implantation procedure was carried out under general or regional anaesthesia with patients in the lithotomy position.

A 16 Ch Foley catheter was placed for guidance in all patients.

For all patients, a traditional surgical technique was used consisting of a perineal incision for cuff placement and inguinal incision for pump unit scrotal placement. After procedure it have been placed into the bladder a 12 Ch Foley catheter. On average it was maintained 2.4 day (1-4 days). Patients were discharged after catheter removal and after they could void. The postoperative patients stayed in the ward for 4.5 day (3–6 days). The device was activated eight weeks later.

Follow-up after implantation

Urine analysis and bladder ultrasonography were conducted to evaluate the residual urine volume, and flow rate measurements were performed after sphincter activation at 1, 3, 6 months after activation and every year up to September 2017.

Patients recorded the number of pads used per day in a 7-day diary before their visits. Total continence was at 0 pads (i.e. completely dry), social continence was at a daily use of 1 pad. Incontinence was given by a daily use of >1 pad.

Success was defined as social continence (0 to 1 pad per day) and an improvement as a decrease in daily pad use (more than a 50% reduction; usually 2 pads per day).

All the patients established the level of pain with a Visual Analogue Scale (VAS) one month after activation of the device. The perioperative complications were classified with Clavien-Dindo questionnaires.

RESULTS

Patient characteristics

From July 2013 to June 2017, 50 patients underwent implantation of a ZSI 375 device in one centre in Poland and were followed up until September 2017. Mean (range) age of 65.5 (24–78) years, with moderate stress incontinence (n = 8) and severe stress urinary incontinence (n = 42). Incontinence was secondary to RP (n = 26), Radiotherapy (n = 2), RP and Radiotherapy (n = 12), TURP (n = 6), Rectal surgery with urinary tract injury (n = 1), pelvic injury (n = 1) and neurological disorder (n = 2) (Table 1).

The mean hospital stay was 4.8 days and the median (range) follow-up was 21.04 (1–50) months.

Complications

Implantation and recovery were uneventful in 48/50 patients (96%). All perioperative patients were classified to grade I Clavien-Dindo classification. Pains related to the presence of artificial urethral sphincter ZSI 375 were assessed with the use of VAS (Visual Analogue Scale). In the pain intensity scale from 0 to 10 patients assessed their ailments as ranging from 0 to 4. Mean

Cent European J Urol

pain intensity value in VAS was 0.82. 35 patients (70%) did not feel any pain resulting from the presence of artificial sphincter. Much more frequently pain occurred in the group of patients with urethral erosion (7/9 patients). No patient experienced bladder hyperactivity after device activation.

Complications leading to revision or permanent device removal arose in 12/50 patients (24%). The recorded complications were urethral erosion (n = 9), infection (n = 0), mechanical failure (n = 3).

All patients with urethral erosion underwent permanent removal of the artificial sphincter.

From the three patients presenting a mechanical failure all of them had a revision and reimplantation with social continence today.

Efficacy

Before device implantation, all patients used ≥ 3 pads for incontinence.

Social continence was achieved in 29/50 patients (58%); continence improved from "4 or more pads" to 2 pads per day in 15/50 patients (30%) (Table 2). Persistent incontinence without complication and with sphincter functioning properly occurred in 6/50 patient (12%).

DISCUSSION

The present study has further limitations: the patients included in the study did have radiotherapy treatment, previous AMS 800 AUS surgery, were operated during surgeons ZSI 375's learning curve and assessment of continence was based on pad numbers and not on pad weight.

The quality of ZSI 375 was improved over the last five years. The reinforced hydraulic circuit of ZSI 375 PF (pre-filled) has been available since March 2015, reducing the operating time should reduce infection and mechanical failure rates.

A better management of pressure increase after activation should help to decrease urethral erosion.

Meanwhile, the two surgeons from the one urological centre included their patients in the study. From their own learning curve in terms of the procedure the present series of implantations, complications and revision rates are similar to the AMS 800 (American Medical Systems, Minnetonka, MN, USA) [20, 21, 22]. In spite of the fact that patient observation time includes the period when operators acquire experience in connection with the learning curve, however the obtained results are comparable to the data from literature concerning the artificial urethral sphincter AMS 800.

In our study efficacy results are similar for the social continence, as evaluated by pad use, was achieved in 29/50 patients (58%). Continence was improved in a further 15/50 patients (30%). Data from the Food and Drug Administration in the USA give a 73% total continence rate for the AMS 800 and an 88% improvement rate [23].

The abdominal approach for reservoir placement can be difficult, especially after robotic surgery and laparoscopy leading to a new approach for the reservoir [24].

A device pre-connected with the absence of an abdominal reservoir simplified the procedure, reducing operating time and risk of infection.

Our median follow-up of 21.04 (1–50) months was long enough to establish the safety and efficacy of the ZSI 375 device and to identify all potential complications. The reported mean time of complication onset for AUS is 19.8 months for erosion, 29.6 months for atrophy, and 68.9 months for mechanical complications [25].

Erosion risk and rate must be discussed and accepted by the patient. From the nine patients presenting erosion, three patients had previous radiotherapy. Radiotherapy is a very well-known

Cent European J Urol

adverse factor for insertion of the sphincter. In other publications urethral erosion ranges from 2% to 17% patients [20, 21, 22, 25, 26].

Infection has not occurred in this study. Postoperative wound infection is one of the most common complication. According to Staerman et al., it can occur in 8.3% of patients, which is comparable to or better than that of AMS 800, Flow Secure or other systems used to treat incontinence in men [4, 5, 12, 15, 20, 21, 22, 26].

Postoperative wound infection is one of the most frequently found complications. According to Staerman and co. it can occur in 8.3% of patients, which is a comparable or better result than in the case of AMS 800, Flow-Secure sphincter, or other systems applied in treating male urinary incontinence.

Three patients had mechanical failures. In all cases of mechanical failure, a new sphincter was implanted with social continence today. In literature mechanical complications are found in about 10% of artificial urethral sphincter implantation cases AUS [22, 26, 27].

A suspicion of one spontaneous reactivation occurred at the same time AMS withdrew its sphincter AMS 800 from the worldwide market because of spontaneous activation in many pumps. Zephyr Surgical Implants add a security system, blocking the sphincter in a safer deactivated position to avoid this kind of complication.

The armed tubing was reinforced in 2011 to avoid kinking making the tubing resistant to breakage.

Usually, surgeons are not fond of preparing AUS, filling the components with a saline solution and removing air bubbles. One sphincter had a revision because of a hydraulic circuit leakage with an unknown cause. ZSI 375 now comes pre-filled, ready to be implanted and the hydraulic circuit is reinforced to reduce the risk of long-term leakage.

CONCLUSIONS

The present study has further limitations: the patients included in the study did have radiotherapy treatment, previous AMS 800 AUS surgery, were operated during surgeons ZSI 375's learning curve and assessment of continence was based on pad numbers and not on pad weight.

Nevertheless, the results of this version of ZSI 375 are similar to AMS 800. This is revealed by comparing our results to the results achieved by other authors presented in the quoted literature. The quality of ZSI 375 was improved over the last five years. The reinforced hydraulic circuit of ZSI 375 PF (pre-filled) has been available since March 2015 reducing the operating time. The ZSI 375 device is safe, effective and the follow-up was long enough to identify all potential complications.

Conflicts of interest

Ireneusz Ostrowski has done one surgical support for Zephyr in Serbia.

Cent European J Urol

Table 1.

Aetiology of incontinence	Infections n (%)	Urethral erosions n (%)	Mechanical complications n (%)
RP (26 patients)	0	4 (15)	2 (8)
RT (2 patients)	0	1 (50)	0
RP+RT (12 patients)	0	2 (17)	0
TURP (6 patients)	0	1 (17)	1 (17)
Rectal surgery (1 patients)	0	0	0
Neurogenic disorder (2 patients)	0	0	0
Pelvic trauma (1 patient)	0	1 (100)	0
Total (50 patients)	0	9 (18)	3 (6)

Table 2.

	Before implantation	12 months after implantation	24 months after implantation	36 months after implantation
Patients, n	50	32	19	15
Pads used/ day, n (%)				
None	0	1 (3)	1 (6)	1 (6.6)
1	0	18 (56)	5 (26)	4 (26.7)
2	0	9 (28)	5 (26)	4 (26.7)
3	7 (14)	0 (0)	3 (16)	4 (26.7)
>= 4	43 (86)	4 (13)	5 (26)	2 (13.3)
Social continence 0,1 pad, n (%)	0	19 (59)	6 (32)	5 (33.3)

Cent European J Urol

Improvement, n (%)	0	10 (31)	9 (47)	6 (40)
Success, n (%)	0	29 (91)	15 (79)	11 (73.3)
Failure, n (%)	0	3 (9)	4 (21)	4 (26.7)



Figure 1. The artificial urinary sphincter ZSI 375.



Figure 2. The implantation of the ZSI 375.

References

1. Scott FB, Bradley WE, Timm GW. Treatment of urinary incontinence by implantable prosthetic sphincter. *Urology* 1973; 1: 252-9.
2. Herschorn S, Bruschini H, Comiter C et al. Surgical treatment of stress incontinence in men. *Neurourol Urodyn* 2010; 29: 179-80.
3. Ostrowski I, Śledź E, Ciecchan J, Golabek T, Bukowczan J, Przydacz M, Wiatr T, Stangel-Wojcikiewicz K, Chłosta PL. Current interventional management of male stress urinary incontinence following urological procedures. *Cent European J Urol*. 2015; 68: 340-347.
4. Montague DK: Artificial urinary sphincter: long- term results and patient satisfaction. *Adv Urol*. 2012; 2012: 835290.
5. Elliott DS and Barrett DM: Mayo Clinic long-term analysis of the functional durability of the AMS 800 artificial urinary sphincter: a review of 323 cases. *J Urol*. 1998; 159: 1206.
6. Brito CG, Mulcahy JJ, Mitchell ME et al. Use of a double cuff AMS800 urinary sphincter for severe stress incontinence. *J Urol*. 1993; 149: 283-5.
7. Guralnick ML, Miller E, Toh KL et al. Transcorporeal artificial urinary sphincter cuff placement in cases requiring revision for erosion and urethral atrophy. *J Urol* 2002; 167: 2075-8.
8. Wilson SK, Delk II Jr, Henry GD, Siegel AL. New surgical technique for sphincter urinary control system using upper transverse scrotal incision. *J Urol* 2003; 169: 261-4.
9. Brian J Flynn, George D Webster. Evaluation and Surgical Management of Intrinsic Sphincter Deficiency after Radical Prostatectomy. *Rev Urol*. 2004 Fall; 6(4): 180-186.
10. Kowalczyk JJ, Spicer DL, Mulcahy JJ. Long-term experience with the double-cuff AMS 800 artificial urinary sphincter. *Urology*. 1996;47:895-897.
11. Kowalczyk JJ, Spicer DL, Mulcahy JJ. Erosion rate of the double cuff AMS 800 artificial urinary sphincter: long-term followup. *J Urol*. 1996;156:1300-1301.
12. Mitra R, de Cógáin and Daniel S. Elliott*. The Impact of an Antibiotic Coating on the Artificial Urinary Sphincter Infection Rate. *J. Urol*. 2013 Jul; 190(1): 113-7.
13. Raj GV, Peterson AC, Toh KL, Webster GD. Outcomes following revision and secondary implantation of the artificial urinary sphincter. *J Urol* 2005; 173: 1242-5.
14. Elliott DS, Barrett DM. Mayo clinic long-term analysis of the functional durability of the AMS800 artificial urinary sphincter: a review of 323 cases. *J Urol* 1998; 159: 1206-8.
15. Lai HH, Hsu EI, Teh BS, Butler EB, Boone TB. 13 years of experience with
16. Artificial urinary sphincter implantation at Baylor College of Medicine. *J Urol* 2007; 177: 1021-5.
17. Wilson SK, Aliotta PJ, Salem EA, Mulcahy JJ. New enhancements of the scrotal one-incision technique for placement of artificial urinary sphincter allow proximal cuff placement. *J Sex Med*. 2010; 7: 3510-5.
18. Sandhu JS, Maschino AC, Vickers AJ. The surgical learning curve for artificial urinary sphincter procedures compared to typical surgeon experience. *Eur Urol* 2011; 60: 1285-90.

Cent European J Urol

19. Frederic Staerman & al: ZSI 375 artificial urinary sphincter for male incontinence: a preliminary study; 2012 BJUI/111, E202-E206.
20. Ostrowski I, Blewniewski M, Neugart F, von Heyden B, Selvaggio O, Iori F, Foley S, Fernández Arjona M, Carvajal Obando A, Pottek T. Multicentre experience with ZSI 375 artificial urinary sphincter for the treatment of stress urinary incontinence in men. *Urologia*. 2017 Aug 1;84(3):148-152.
21. Hajivassiliou CA. A review of the complications and results of implantation of the AMS artificial urinary sphincter. *Eur Urol* 1999; 35: 36-44.
22. Venn SN, Greenwell TJ, Mundy AR. The long-term outcome of artificial urinary sphincters. *J Urol*. 2000; 164: 702-6.
23. Kim SP, Sarmast Z, Daignault S, Faerber GJ, McGuire EJ, Latini JM. Long-term durability and functional outcomes among patients with artificial urinary sphincters: a 10-year retrospective review from the University of Michigan. *J Urol* 2008; 179: 1912-6.
24. Bauer RM, Gozzi C, Hübner W et al. Contemporary management of postprostatectomy incontinence. *Eur Urol* 2011; 59: 985-96.
25. Wilson DK, Delk IJR. Ectopic placement of AMS 800 urinary control system pressure-regulating balloon. *Urology* 2005; 65: 167-70.
26. Lai HH, Hsu EI, The BS, Butler EB, Boone TN. 13 years of experience with artificial urinary sphincter implantation at Baylor College of Medicine. *J Urol* 2007; 177: 1021-5.

Cent European J Urol

27. Staerman F, G-Llorens C, Leon P, Leclerc Y. ZSI 375 artificial urinary sphincter for male urinary incontinence: a preliminary study. *BJU Int.* 2013; 111(4 Pt B): E202-6.
28. Herschorn S, Thuroff J, Bruschini H, et al. Surgical treatment of urinary incontinance in men. In: *Incontinance. Third International Consultation.* Abrams P, Cardozo L, Khoury S, Wein A (red.). Health Publications Ltd, Paris 2005; 1241-96.